## **Specifications**

For other materials or modifications, please consult TESCOM.

### **OPERATING PARAMETERS**

Pressure rating per criteria of ANSI/ASME B31.3

### **Maximum Inlet Pressure**

4500 psig / 310 bar

### **Outlet Pressure Ranges**

50 mm Hg absolute - 15 psig / 1.0 bar 50 mm Hg absolute - 50 psig / 3.4 bar 50 mm Hg absolute - 100 psig / 6.9 bar 50 mm Hg absolute - 350 psig / 24.1 bar

### **Design Proof Pressure**

150% maximum rated

#### Leakage

Internal, Bubble-tight

## Operating Temperature<sup>1</sup>

-15°F to 140°F / -25°C to 60°C

## Flow Capacity

 $C_{\rm V} = 0.06$ 

## **Maximum Operating Torque**

30 in-lbs / 3.4 N • m

### MEDIA CONTACT MATERIALS

## **Body**

Brass or Nickel Plated Aluminum

## Diaphragm

Buna-N, Ethylene Propylene, or Viton®

### Seat

Teflon® (Inlet 500 psiq / 34.5 bar maximum), CTFE, or Vespel®

## Friction Sleeve (inner)

Teflon®

## Friction Sleeve (outer)

316 Stainless Steel

## Filter (40 micron)

316 Stainless Steel

## **Remaining Parts**

300 Series Stainless Steel

## **OTHER**

### Cleaning

CGA 4.1 and ASTM G93

## Weight (without gauges)

**Brass:** 2.4 lbs / 1.1 kg **Aluminum:** 1 lb / 0.5 kg

1. For extended temperatures from -40°F to  $400^{\circ}F/-40^{\circ}C$  to  $204^{\circ}C$ , consult Tescom. Viton®, Teflon®, and Vespel® are registered trademarks of E.I. du Pont de Nemours and Company.



TESCOM DA Series absolute pressure reducing regulator is designed with an elastomeric diaphragm and provides accuracy to +/- 0.1 psig / 0.007 bar. This regulator may be used to reduce pressure from a supply source up to 4500 psig / 310 bar into a vacuum environment.

## **Applications**

- · Instrumentation testing
- · Calibration equipment

### **Features and Benefits**

- Compact in size and highly sensitive
- Economical
- Quick response and accurate diaphragmtype regulation
- Excellent repeatability
- Non-venting
- Low operating handknob torque
- Captured bonnet is available
- Panel mounting option is available
- Variety of porting options, body materials and soft goods








# **TESCOM**

# **DA Series Regulator Drawing**








# **DA Series Regulator Flow Chart**

For more information on how to read flow curves, please refer to the Flow Curves and Calculations document (debul2007x012) in the TESCOM catalog or on www.tescom.com.





## **DA Series Regulator Part Number Selector**

Repair Kits, Accessories & Modifications may be available for this product. Please contact TESCOM for more information.

Example for selecting a part number:

DA	1	3	0	)	K	1	В	9	Α	FF9
BASIC SERIES		BODY MATERIAL	OUTLET PRESSURE RANGES		SEAT MATERIAL	FLOW C <sub>V</sub>	DIAPHRAGM MATERIAL	MOUNTING	STANDARD PORTING	INLET, OUTLET AND GAUGE PORT TYPE
			STANDARD	ABSOLUTE <sup>1</sup>		<b>~</b> V	WIN LINK		CONFIGURATION	AND SIZE
DA	Standard Pressure Reducing 1 - Handknob adjust 2 - Screwdiver adjust 3 - Captured bonnet Hand adjust 4 - Captured bonnet Screw adjust Absolute Pressure Reducing A - Handknob adjust B - Screwdriver adjust C - Captured bonnet Hand adjust D - Captured bonnet Screw adjust E - Dome loaded (maximum dome pressure 125 psig / 8.6 bar)	1 – Brass 3 – Aluminum	<ul> <li>0 – N/A</li> <li>1 – 0-15 psig 0-1.0 bar</li> <li>2 – 0-50 psig 0-3.4 bar</li> <li>3 – 0-100 psig 0-6.9 bar</li> <li>4 – 0-350 psig 0-24.1 bar</li> </ul>	50 mm Hg absolute - 100 psig / 6.9 bar (Dome loaded only) 50 mm Hg absolute - 15 psig / 1.0 bar 50 mm Hg absolute - 50 psig / 3.4 bar 50 mm Hg absolute - 100 psig / 6.9 bar 50 mm Hg absolute - 1100 psig / 6.9 bar 50 mm Hg absolute - 1100 psig / 6.9 bar	K - CTFE Inlet 4500 psig / 310 bar maximum V - Vespel® Inlet 4500 psig / 310 bar maximum T - Teflon® Inlet 500 psig / 34.5 bar maximum	1 - C <sub>V</sub> = 0.06	B – Buna-N E – Ethylene Propylene V – Viton®		A - No gauge ports  D - Out gauge port at 90°  B - In and Out gauge port at 60°  L - In and out gauge port at 90°  L - In and out gauge port at 90°	B – 1/4° SAE F – 1/4° NPTF J – 1/4° MS33649 9 – None

 $\Lambda$ 

WARNING! Do not attempt to select, install, use or maintain this product until you have read and fully understood the TESCOM Safety, Installation and Operation Precautions.

